

OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, KEONJHAR

No 4577

Dated 23 / 09 / 2019

Terms and Conditions of Tender submission & Finalisation

Sealed Tenders are invited from the interested firms / agencies / organization for supply of Baby weighing scale with Sling and Clinical Digital Thermometer to CDM&PHO, Keonjhar as per specification detailed in **Annexure "A"**

Last date for submission of Tender	14 . 10 .2019 Time 05.30 PM Address: CDM&PHO Keonjhar At/Po/Dist – Keonjhar, PIN – 757001 Trough Speed Post / Regd Post only
Date, Time & Place of Opening of Tender	15 . 10 .2019 Time 11 AM At – Office Chamber of the CDM&PHO, Keonjhar
Documents to be submitted	a) Copy of up-to-date valid GST number Clearance Certificate b) Copy of PAN Card c) Copy of Registration of Agency / Organization d) Annual Average turn over 10 Lakhs or above for the last three years(2015-16,2016-17,2017-18,) in this regard a certificate to be obtained from chartered accountant. e) Audited financial statement (Receipt & Payment Account, income and expenditure account, Balance Sheet) for last 3 years (2015-16,2016-17,2017-18,) f) Three years experience (Proof of at least 5 work orders of hospital supply) g) Affidavit of Non-black listed firm for execution of Govt similar works.(From Executive Magistrate / Notary) h) Undertaking for supply of the materials as per the specifications only. i) Manufacturer authorization certificate The previous non-performing agency or organization will be disqualified.
Signing of Documents	All papers / documents submitted with the Tender must be signed by the Bidder
EMD and Price for Bidding Documents	` 500/- (which is non refundable)in form of Bank Draft drawn in favour of ZSS,NRHM, Keonjhar towards Price for Bidding Documents. ` 2000/- in form of Bank Draft in favour of ZSS,NRHM Keonjhar towards EMD. EMD will be refunded to the unsuccessful bidders after finalisation of Tender Process. The EMD of successful bidder will be refunded after completion of the supply.
Completion of the Process	Seven days from the placement of Supply Order in each case.
Submission of Bid Documents	The Bidders must submit the Technical and Financial Bid Separately in the sealed envelope and both the Technical & Financial Bid Envelopes must be kept in another envelopes (Outer). The outer envelope must be superscribed on the top of the envelope " Tender for supply of "ASHA HBNC KIT" and to be addressed to the CDM&PHO,Keonjhar At / PO/ Dist – Keonjhar


Chief District Medical & Public Health Officer
Keonjhar

Validity	The rate must be quoted as per specification, term & condition and will be valid for one year.
Sample	The Bidder must submit the sample of the each items as per the specifications only.
The undersign reserves all the rights to cancel or reject the any or all tenders without assigning any reason thereof. The decision of the committee will be treated as final.	

Annexure "A"

List of items with specifications

SI	Name of the items	Specifications
1	Baby weighing Scale with Sling	<p>Should have touch Nylon plastic(Black) body with demonstration photo in white colour, Maximum Load capacity is 5 Kg, Each product should be stamped by Weight & Measurement department. Should have colour coded Reading scale. Should have zero adjustment, should have inbuilt overload protection, Minimum graduation should not be less than 100gms, Scale should be fitted with an approx 3" rod handle with non-corrosive metallic ring on top and "S" shaped hook at the bottom for suspending the sling bag, Spring used in the scale is made up of heavy duty steel sling non corrosive grade II.</p> <p>Sling bag should be made up of double stitched touch & tear resistant 100% polyester to weigh the baby (with load carrying capacity of at least 7 Kgs) Sling bag width at the middle of the .24". Two numbers sling bag should be provided in each baby weighing scale.</p>
2	Clinical Digital Thermometer	<p>Digital Thermometer with Fahrenheit scale. Safe to use automatic, No Glass No Mercury. Measurement range 90.0° F to 109° F. Measurement accuracy : +/- 0.2° F between 90.0° F to 102° F +/- 0.3° F between 96.0° F to 98° F +/- 0.3° F between 102° F to 106° F (Type testing certificate from lab relating to measurement accuracy and need to be furnished in technical bid)</p> <p>Should have auto shut off facility. Display should be LCD type. Beeper alarm of approx .10 seconds sound signal when peak temperature reached. Average measurement time Oral: 30 Sec, Rect: 30 Sec, Auxiliary 60 sec, Battery operated and should have more than 20 hours of battery life.</p> <p>Warranty: 1 year</p>


 Chief District Medical & Public Health Officer
 Keonjhar

Technical Bid Document for supply of HBNC Kit

Name of the firm / agency (Proof of Registration to be Submitted)	
Address of the Firm / Agency	
Name of the Contact person	
Email of the contact person	
Contact Number	
PAN Card No of the organization (Firm/Agency/Proprietor). Copy to be attached	
Copy of Up-to-date GST Certificate and GST No.	
Average Annual Turnover of Firm/Agency/Organization of 10 lakhs or above for the last 3 years(proofs to be attached) (2015-16,2016-17,2017-18,) in this regard a certificate to be obtained from chartered accountant.	
Audited financial statement (Receipt & Payment Account, income and expenditure account, Balance Sheet) for last 3 years (2015-16,2016-17,2017-18,)	
Affidavit (Notary/Executive Magistrate) of non-black listed Firms for execution of Govt similar works.As per Annexure:B	
Price for bidding document ` 500/- (which is non refundable) in the form of Bank Draft in favour of ZSS,NRHM , Keonjhar payable at Keonjhar and EMD of ` 2000/- (Which is refundable) in Form of Bank Draft in favour of ZSS,NRHM Keonjhar.	
Sample as per specification	
Manufacturer authorization certificate	
Undertaking for supply of the materials as per the specifications only	
Three years experience (Proof of at least 5 work orders of hospital supply)	

(Annexure "B")

DECLARATION

I Mr / Mrs _____ on behalf of

_____ (Firm/Agency/Organization Name) declared that I

/ We are not blacklisted by any Central / State Govt. / Public Sector undertaking in India. I have given consent that the supply of above materials will be done in the stipulated time as per given specification. I confirm that the information that I have provided above is true & correct.


Date:

Signature

Place:

Name

Designation:


 Chief District Medical & Public Health Officer ,
 Keonjhar

FINANCIAL BID for supply of Baby weighing scale with sling & Clinical Digital Thermometer

SI	Name of items as per specification	Specifications	Unit	Base Price (Excludi nd tax)	Tax	Rate in Rs. (Inclusive of all Tax &Transportation)
1	Baby weighing Scale with Sling	Should have touch Nylon plastic(Black) body with demonstration photo in white colour, Maximum Load capacity is 5 Kg, Each product should be stamped by Weight &Measurement department. Should have colour coded Reading scale. Should have zero adjustment, should have inbuilt overload protection, Minimum graduation should not be less that 100gms, Scahle should be filled with an approx 3" rod handle with non -corrohsive metallic ring on top and "S" shaped hook at the bottom for suspending the sling bag, Spring used in the scale is made up of heavy duty steel sling non corrosive grade II. Sling bag should be made up of double stitched touch & tear resistant 100% polyester to weigh the baby(with load carrying capacity of at least 7 Kgs) Sling bag width at the middle of the :24".Two numbers sling bag should be provided in each baby weighing scale. "	Per Piece			
2	Clinical Digital Thermometer	Digital Thermometer with Fahrenheit scale. Safe to use automatic, No Glass No Mercury. Measurement range				


 Chief District Medical & Public Health Officer
 Keonjhar

	<p>90.0° F to 109° F. Measurement accuracy : +/- 0.2°F between 90.0° F to 102° F : +/- 0.3°F between 96.0° F to 98° F : +/- 0.3°F between 102° F to 106° F (Type testing certificate from lab relating to measurement accuracy and need to furnished in technical bid) Should have auto shut off facility. Display should be LCD type. Beeper alarm of approx .10 seconds sound signal when peak temperature reached. Average measurement time Oral : 30 Sec, Rec: 30 Sec, Auxiliary 60 sec, , Battery operated and should have more than 20 hours of battery life. Warranty: 1 year</p>				
--	--	--	--	--	--

Date:

Place:

Signature

Name

Designation: Chief District Medical & Public Health Officer
Keonjhar

LO
23/9/19