

कार्यालय मुख्य चिकित्सा अधिकारी, उधम सिंह नगर


दूरभाष नं०-05944-246590, ई मेल- cmousn2018@gmail.com

पत्रांक:-Covid-19/2020-21/ 3537

दिनांक:-24.06.2020

संशोधित-उपकरण अल्पकालीन-निविदा सूचना

इस कार्यालय के पत्र संख्या- पत्रांक:-निविदा/उपकरण/20-21/3410 दिनांक 19.06.2020 द्वारा प्रकाशित निविदा में उपकरणों के Specification के पृष्ठ संख्या 13 में आंशिक संशोधन किया गया है, जिसे इस कार्यालय की वेब साइट वेबसाइट- www.cmousnagar.com & www.usnagar.nic.in पर देखा जा सकता है। निविदा की शेष शर्तें यथावत रहेंगी।


मुख्य चिकित्सा अधिकारी
उधम सिंह नगर।

कार्यालय मुख्य चिकित्सा अधिकारी, उधम सिंह नगर

पत्रांक:-Covid-19/2020-21/ 3538

दिनांक:-24.06.2020


सेवा में

जिला सूचना अधिकारी,
उधमसिंह नगर

विशय- अल्प कालीन उपकरण निविदा दैनिक समाचार पत्र के उत्तराखण्ड संस्करण में प्रकाशित करने के सन्दर्भ में।
महोदय

उपरोक्त विशयक अवगत कराना है, कि संलग्न उपकरण निविदा को 03 दैनिक समाचार पत्रों में दिनांक 25 जून 2020 को प्रकाशित होने वाले किन्ही 3 समाचार पत्रों में न्यूनतम दरों पर एक बार प्रकाशित करने का कष्ट करें।

भवदीय


मुख्य चिकित्सा अधिकारी
उधमसिंह नगर।

S. NO.	NAME OF EQUIPMENT	Specification
1	Blood Cell Counter	<ul style="list-style-type: none"> • 22 Reportable parameters including P-LCC, P-LCR • Electrical impedance method for Cell count • Cyanide free reagent for Haemoglobin test • 9 µl sample volume • Large coloured LCD touch screen of 10.4 inch • Throughput 60 samples/ hr • 3 counting modes -Whole blood, Capillary and pre-dilute • Large storage capacity of 50000 samples
2	Heamoglobin Meter	<ul style="list-style-type: none"> *SAMPLE VOLUME:10µL *MEASURING TIME<15SEC. *BATTERY LIFE:1000TEST *MEMORY :1000TEST RESULT *CALIBRATION :AUTO CALIBRATION BY CODE CHIP MEASURING RANGE:5-25GM/DL Aspen Diagnostic (P) ltd.
3	Endo Tracheal Tube Adult/Pedia	Standarded Quality
4	Blower/ Room Heater	Standarded Quality ISI Brand,
5	LSCS Set	Standarded Quality ISI Brand,
6	Surgical Diathermy Cut/ Cutery	<ul style="list-style-type: none"> • Electro surgical unit with Cut/Coag/Bipolar facility. • Two independent output activation in monopolar mode. • ISO & CE certified. • Audio visual alarms for various application. • Warranty : 3 years
7	Self Retractor	Standarded Quality ISI Brand,
8	Urology Set	Standarded Quality ISI Brand,
9	Harmonic Scissor for Laproscopic	Standarded Quality ISI Brand
10	Reagent Based Glucometer	Standarded Quality ISI Brand


 Chief Medical Officer
 Udham Singh Nagar,