

 महाराष्ट्र शासन जिल्हा रुग्णालय, उस्मानाबाद 		
कोड क्रमांक : - दूरध्वनी क्रमांक (वैयक्तिक): - कार्यालय दूरध्वनी क्रमांक १ : - NCD विभागातील क्रमांक. -	०२४७२-२२२६५० २२६९२४ २२७००५ २२६७०५	पत्ता : -जिल्हा शल्यचिकित्सक,जिल्हा रुग्णालय, उस्मानाबाद, कार्यालय मारवाडी गल्ली,उस्मानाबाद फॅक्स नंबर: -०२४७२ - २२६९२४ Web Site : -http://www.maha-arogya.gov.in ई-मेल :-cs_osmanabad1@rediffmail.com
आरोग्य सेवा		जाक्र/जिरुउबाद/एनएचएमईपीएफएनसीडी/५७४०/२०२० दिनांक :- 27.2.2020

दरपत्रक सुचना

उपरोक्त दरपत्रक सुचनेनुसार आपणास कळविण्यात येते की, जिल्हा रुग्णालय उस्मानाबाद येथे राष्ट्रीय मौखिक आरोग्य कार्यक्रम चे साहित्य सामुग्री खरेदी करावयाचे आहे. तरी त्यांचे खालील नमुन्यानुसार या कार्यालयास ७ दिवसाच्या आत दर सादर करण्यात यावे.

Name & Description of Item

Pits & Fissure Selant Consumable

Specification-

1) Glass inorner based with highest Fluride release in auto cure mode

Packing - 15 gm powder,10 g liquid,6g DENTIN

CONDITIONAR or 12.5gm powder,8.5 mi liquid.

Avilable in coloured (Self cure) and white (self cure)

Mesureing Scoop & mixing Pad

Provide Instruction Manual

2) Release high fluoride ions

3) Should be Radiopaque

4) Very high adhesion to dentin & enamel & high abrasion resistance

5) Easy t manipulate

6) Should facilate remineralization

7) The Product should have internationally published scientific studies supporting its used for the intended purpose

8) Having high wettability good flow into narrow pits & Fissures

9) should supply in pkts with manufacturer name description of the instrument & Lot No.

10) Product should sealed at the time of Demonstration & Supplier should provide sample for clinical trial if asked for by technical committee

11) Should by CE Europien or US FDA certified

12) The vendor should confirm the packing is Original

13) Vendor should provide guaranty towards replacement to ultimate consignee against manufacturing defect whenever detected & also agree to exchange the sealed package before the expiery of self life/if expired unused

