

समाहरणालय, कैमूर (भभुआ)

(जिला योजना कार्यालय)

अल्पकालीन निविदा आमंत्रण सूचना

कैमूर जिलान्तर्गत संचालित विभिन्न योजनाओं यथा 15वीं वित्त आयोग, मुख्यमंत्री क्षेत्र विकास योजना इत्यादि के तहत कोविड महामारी के रोकथाम संबंधी चिकित्सा एवं अन्य सामग्री/उपकरणों की अधिप्राप्ति हेतु ख्याति-प्राप्त अधिकृत/रजिस्टर्ड उत्पादकों/विक्रेताओं से मुहरबंद निविदा आमंत्रित की जाती है। वित्तीय/तकनीकी निविदा अलग-अलग मुहरबंद लिफाफे में दिनांक 29.06.2021 को 04:00 बजे अपराह्न तक जिला योजना कार्यालय, कैमूर (भभुआ) में रजिस्टर्ड डाक/कुरियर या स्वयं अथवा प्रतिनिधि के माध्यम से दाखिल कर सकते हैं। उप विकास आयुक्त कैमूर (भभुआ) की अध्यक्षता में क्रय समिति के समक्ष प्राप्त निविदा का निस्तारण उनके कार्यालय प्रकोष्ठ में दिनांक-30.06.2021 को 11:00 बजे पूर्वाह्न में की जायेगी, जिसमें निविदादाता या उनके प्राधिकृत प्रतिनिधि भाग ले सकते हैं। तकनीकी एवं वित्तीय निविदा अलग-अलग लिफाफे में जमा करना होगा एवं दोनों लिफाफे पर तकनीकी एवं वित्तीय निविदा, मोटे अक्षरों में लिखा होना चाहिए तथा मुहरबंद रहना चाहिए। तकनीकी निविदा में सफल निविदादाता का ही वित्तीय निविदा खोला जाएगा। निविदा में निर्धारित दर पर ही जिला प्रशासन द्वारा आवश्यकतानुसार संबंधित सामग्रियों/उपकरणों की क्रय की जाएगी।

➤ तकनीकी निविदा की शर्त:-

A bidder participating in the procurement process must possess the following minimum & mandatory eligibility criteria, failing which the concerned bid shall be rejected. All necessary documents must be "Self-Attested" and enclosed in the technical envelope.

S. No.	Basic Requirement	Specific Requirements	Documents Required
1.	Bidder's Legal Entity (Joint Venture or Consortium or Association of bidders shall not be allowed)	The bidder should be duly registered from 5years before, Under either the Bihar shops & Establishments Act, 1953. OR A company registered under Indian Companies Act, 1956. OR A partnership firm registered under Indian Partnership Act, 1932.	i. Registration Certificate of shop & Establishment from Last 5 years (in case of proprietorship firm) ii. Registration certificate and Memorandum of Association issued by Registrar of Companies in case of a registered company and in case of another statutory or registered body, certificate of incorporation or registration issued by concerned authority. Power of attorney in favour of the person signing the bid. (in case of company registered under company Act). iii. Partnership Deed and valid registration certificate with the Registrar of Firms in case of partnership Firms. Power of Attorney in favour of the partner signing the bid, authorizing him to represent all partners of the firm (in case of partnership firm)

2.	Financial Turnover	Average annual turnover of the bidder during financial year 2019-2020 should be Minimum 5 Crore.	Audited balance sheet of relevant FYS. + CA Certificate with CA's Registration Number/ Seal
3.	Financial Net Worth	The net worth of the bidder should be minimum, 1 Crore as on 31-03-2020.	CA Certificate with CA's Registration Number/Seal
4.	Tax Registration and Clearance	Firms must have registration in Bihar GST. The bidder should have a registered number of GST and Income Tax/PAN number.	Copies of relevant certificates of registration.
5.	GST Return	Firm must have to submit proof of GST return of minimum Rs.50 Lakh as on 31-03-2021.	Copies of GSTR1 or GSTR3B
6.	EMD (Security Deposit)	Demand draft of INR. 2,00,000/- in favour of "Deputy Development commissioner, Kaimur (Bhabua)"	Demand draft should be issued from a Nationalized Bank
7.	Character Certificate	Firm should submit the character certificate of proprietor (in case of proprietorship firm) or Character certificate of Director's who is signing the bid (in case of Company Registered under Company Act) Or Character certificate of partner, having right to sign the bid (in case of partnership firm)	Copy of character certificate
8.	Mandatory Undertaking	Bidder should: - a) Not be insolvent, in receivership, bankrupt or being wound up, not have its affairs administered by a court or a judicial officer, not have its business activities suspended and must not be subject to legal proceedings for any of the above reasons; b) Not have, and their directors and officers not have, been convicted of any criminal offence related to their	Affidavit/Self Declaration

		<p>professional conduct or the making of false statements or misrepresentations as to their qualifications to enter into a procurement/ bid contract within a period of three years preceding the commencement of the procurement/bidding process, or not have been otherwise disqualified pursuant to debarment proceedings;</p> <ul style="list-style-type: none">c) Not have been black-listed by any government or any government statutory agency.d) Not sublet the contract, if awarded.e) Have submitted only one bid.	
--	--	---	--

Financial Bid Format

(To be printed on company letter head)

Sl. No.	Particulars/ Items	Specifications
1	General Hospital Bed.	<p>Overall Size : 206 L X 90 W X 60 H Cms. , Frame work made Rectangular M.S. Tube. , Top of round perforated CRC M.S. Sheet. , Tubular head & foot bows of unequal height with vertical tube Support. , Legs fitted with PVC stumps. , Finish Epoxy Powder Coated.</p> <ol style="list-style-type: none"> 1. Should have four sections. Top flat platform should be made of perforated CRC sheet of thickness of 18G or better. 2. Bed frame must be sturdy and stable to support weight of at least 170kg. The frame structure should be made up of at least 16 G CRC, rectangular / circular pipe of 100 mm x 30 mm. 3. Bed frame mounted on trolley base made up of 100mmx30mm CRC rectangular pipe of 16 gauge. 4. All adjustments for fowler position must be obtained from crank shaft, manually operated with Stainless Steel / ABS foldable handle on both the shaft. 5. The finished bed must be rust proof, pretreated and powder coated with 7 tank chemical process. 6. The bed should have telescopic side rails of 202 Grade stainless steel of 20 Gauge with spring loaded locking arrangement on both sides. 7. Should have easily removable head and foot panels made up of ABS plastic. 8. Fowler bed should be of following dimension: Mattress area of Length 2000 to 2010 mm X Width 900 to 1000mm Height: -500 to 550mm (without mattress). 9. Should have strong & good quality single wheeled total locking type Swivel Castors of 125 to 150 mm diameter with breaks on all four castors for stabilized position. (GI Castors) 10. There should be suitable buffer mechanism to avoid hitting of the bed to the wall from all sides. 11. Should have provision of fixing suitable rod for hanging intravenous / irrigation fluid bottle on both sides at head end and foot end. Each bed should be supplied with 2 no. good quality I. V. rod Stainless Steel 202 Grade. 12. Should have hooks on bed frame on both side for holding urine / drainage bag (at least 2 Nos. on each side). <p>Mattress to be four sectional and to be combination of good quality foam and coir layer. Dimensions: - Exactly fitting to the Bed, 4 – inch thickness.</p>
	Pediatric Bed	
	Hospital Fowler Bed.	
2	Locker	Dimension - 16" L X 16" W X 32" H. . Four MS legs fitted with PVC Stumps. , Box and door made of MS sheet. , S.S. Top made by machine pressed joint less/bend pressed with three side raise edges., Finish : Epoxy Powder coated.
3	Saline Stand	5 Pronged base fitted with 5 cm dia castors. 10 mm C.P. Adjustable double hook I.V. Rod to hold I.V. rod at one fixed height by knob. Powder Coated base with Powder Coated pipe and Chrome plated rod.
4	Oxygen Cylinder (Jumbo)	<p>Water Capacity (Liters) D- Type(46.7 Liter)</p> <p>Working Pressure 150 kg/cm² at 15oc</p> <p>Type of Cylinder Filled</p> <p>Gas Capacity 7000 Liters</p> <p>Cylinder Material 34 Magna Steel</p>
5	Oxygen Cylinder B type(Small)	<p>Water Capacity (Liters) B-Type (10 Liter)</p> <p>Usage Medical Use</p>

		<p>Working Pressure 150 kgf/cm²</p> <p>Type of Cylinder Filled</p> <p>Gas Capacity 1500 Liters</p>
6	Oxygen Flow Meter	<p>Metering tube and cover made of Unbreakable Poly carbonate</p> <p>Flow adjustment by Needle Valve equipped with inlet filter- 100um</p> <p>Flow rate range 0- 15 liters per minute</p> <p>Flush Flow 60 liters per minute</p> <p>Read the flow on the center of the ball</p> <p>Inlet pressure 60 psi</p>
7	Oxygen Concentrator	<p>Flowrate: O-SL/min, Purity:93%±3% Outlet Pressure (Mpa): 0.04-0.07</p> <p>Sound level(dB):<50</p> <p>Power:AC220V±10%,50/60±1Hz; Power:<480W</p> <p>Timer: accumulating timing; present timing</p> <p>Alarm: Power failure, High & Low pressure, High temperature</p> <p>Net Weight: 25kg</p> <p>Size(mm):365x375x600</p>
8	Stretcher Trolley	<p>Overall Approx. Size : 84" L X 22" W X 32" H. , Frame work made of 1.25" X 16 G vertical and 1" X 18 G horizontal CRC Tubes. , Trolley mounted on 6" dia castors 2 with brake. , Removable stretcher top made of MS sheet with S.S. Handles. , Finish : Pre-treated and Epoxy Powder Coated.</p>
9	Wheel Chair	<p>16" Seat with Fixed Armrest Front with 4" Solid Castors & Rear 26" Solid Wheels with brakes. , Size 39.5" L X 28" W X 34.5 " H.</p>
10	Patient Examination Table	<p>183 L X 60 W X 81 H cms. , Frame work of CRC tubes. , Legs fitted with PVC stumps. , 5 cms foam cushioned top covered with rexine. , Pre-treated and epoxy powder coated finish.</p>
11	Pediatric Cuff	<p>As per standard</p>
12	Nebulizer	<p>Method of Nebulization Piston Driven Power 120VAC, 60Hz</p> <p>Operating Temperature Range 50°F TO 104°F (10°C TO 40°C)</p> <p>Operating Humidity Range 10% to 90% RH non-condensing</p> <p>Storage Temperature Range 5°F TO 122°F (-15°C TO 50°C)</p> <p>Storage Humidity Range 10% to 90% RH non-condensing</p> <p>Weight 4.2 lb.</p> <p>Width 6.8"</p> <p>Height 4.2"</p> <p>Length 11.7"</p> <p>Max Pressure (Compressor) 30 psi</p> <p>Flow Range (Compressor) 0 - 14 liters/min</p> <p>Particle Size (Nebulizer) 0.5 to 5 microns</p> <p>Capacity (Nebulizer) 5 ml</p>
13	Pulse Oximeter	<p>Display format: LCD Back light: LCD</p> <ul style="list-style-type: none"> > SP02:35-99% > Pulse ratio:30-250BPM > Pulse wave: LCD graphic display > Battery voltage low indication: LCD Signals. > Power requirement: two MA size 1.5 Alkali Battery.
14	Suction Machine	<ul style="list-style-type: none"> <input type="checkbox"/> Noise level of suction apparatus 50 db.±-03 db. <input type="checkbox"/> Leakage current of suction units less than 84 UA. <input type="checkbox"/> Electrical requirement - 220 v, 50Hz, 1 phase. <input type="checkbox"/> Rotary vane having flat spring vacuum pump with -740 +10mm Hg. <input type="checkbox"/> Free air displacement 40 ltrs. / min. <input type="checkbox"/> Heavy duty HN -65 antistatic castors, 320-watt, 1440 RPM, 0.50 H. P. Electric Motor <input type="checkbox"/> 100 \mm vacuum Gauge.

		<input type="checkbox"/> Non collapsible silicon tubing <input type="checkbox"/> 2 x2.0 Lts. Autoclavable Polycarbonate jars having autoclavable Lids with overflow safety. <input type="checkbox"/> Bacterial filter fitted on top.
15	Three Seater Chair	Perforated Sheet. , Galvanized arm rest
16	Multi Para Monitor	<ul style="list-style-type: none"> • 5 standard parameters: ECG, RESP, NIBP, SP02.1-TEMP • Real time S-T segment analysis and ARR analysis. • Multi-display selectable including standard, large font, trend Oxy CRG dynamic • Multi 7-lead ECG waveforms display in phase and capture dynamic waveforms. • Large volume of tabular and graphic trends information storage and easy to recall • Efficient resistance to interference of Defibrillator.
17	Oxygen Mask	<ul style="list-style-type: none"> • Non-toxic & non-irritant material. • Shall have elastic strip to hold mask in place. • Mask anatomically contoured for comfort. • Swivel connector for convenient attachment of oxygen tubing. • Tube should not kink. • Tube should be highly flexible to ensure continuous flow of oxygen. • Adult and pediatric sizes. • Compatible with any standard Nebulizer.
18	Pediatric Mask	<ul style="list-style-type: none"> • Non-toxic & non-irritant material. • Shall have elastic strip to hold mask in place. • Mask anatomically contoured for comfort. • Swivel connector for convenient attachment of oxygen tubing. • Tube should not kink. • Tube should be highly flexible to ensure continuous flow of oxygen. • Pediatric sizes. • Compatible with any standard Nebulizer.
19	Thermometer	<ul style="list-style-type: none"> • Digital with beep
20	Baby Warmer	<ul style="list-style-type: none"> • This manually controlled Infant Warmer used for supplying supplementary heat to the newborn neonate. • It features a standard heater head, dovetail rail accessory mounting system, and an integrated bed. • Specification: Micro Processor Controlled with Servo & Manual modes of operation. • This infant warmer provides automatic over heat and over temperature protection in both manual and servo modes Quartz Infrared heater element with very effective heat Radiation. • Temperature Probe ensures an accuracy of +/- 0.2 Deg. C. • A probe is inter-changeable without calibration. • An integral Bed with up-down tilting facility Optimum Bed dimensions ensure optimum workspace for the nursing staff Display of set temperature and infant temperature. • Comprehensive Alarms and safety cut-offs. • Wide base for stability and Castors with brakes for good mobility. • Easy to use and clean Convenient Storage tray for keeping essential things • Halogen Lamp for observations.
21	Emergency Trolley	<ul style="list-style-type: none"> • Two section top. • Height adjusted by a foot operated imported hydraulic pump from 660 mm to 910 mm. • Gas spring assisted Trendelenburg/Reverse Trendelenburg positions. • Complete with corner buffers, synthetic rubber covered handles, sliding X-ray cassette, accessories tray, oxygen cylinder holder, SS telescopic IV rod and swing away SS side rails.

		<ul style="list-style-type: none"> • Pretreated and powder coated. • Four imported swivel castors, 15cms dia, two with total lock X-ray Permeable removable stretcher, backrest raised on the ratchet.
22	Pediatric Drip Set	<ul style="list-style-type: none"> • As per standard
23	Child Stethoscope	<ul style="list-style-type: none"> • single-sided Chest Piece • For Paediatrician • Single Lumen Tube
24	Phototherapy Machine	<ul style="list-style-type: none"> • Double Surface
25	Pediatric Nebulizer	<ul style="list-style-type: none"> • Respiratory Steam Nebulizer Machine with Complete Kit for Baby, Adults, kids & Asthma Inhaler Patients • 0.2 ml/min
26	Surgical Mask	<ul style="list-style-type: none"> • Disposable 3 ply protective mask with adjustable nose bridge and ear loop. • Bacterial filtration efficiency (BFE) - 99.9%.
27	Gloves 6½ No.	<ul style="list-style-type: none"> • Material: shall be made by the dipping process using Natural Rubber Latex. Finish of outer surface shall be smooth. The gloves shall be transparent or translucent. • Lightly powdered with absorbable dusting powder. • Tensile strength: Min 18MPa & Elongation at break: Min 700%. • The gloves shall be free from visible imperfection. • The gloves shall not contain any ingredients known to be harmful either to the wearer or to persons they may come in contact with.
28	Gloves 7 No.	<ul style="list-style-type: none"> • Material: shall be made by the dipping process using Natural Rubber Latex. Finish of outer surface shall be smooth. The gloves shall be transparent or translucent. • Lightly powdered with absorbable dusting powder. • Tensile strength: Min 18MPa & Elongation at break: Min 700%. • The gloves shall be free from visible imperfection. • The gloves shall not contain any ingredients known to be harmful either to the wearer or to persons they may come in contact with.
29	Sanitizer	<ul style="list-style-type: none"> • 5 Litre, 500 ml. • Ethanol 80% v/v, Glycerol 1.45% v/v, Water 18.45% v/v. • Preferred Ayurvedic ingredients.
30	O T Table Hydraulic	<p>Five sectional stainless steel top which gives easy various surgical positions. , UP & DOWN by Hydraulic System. , Standard Positioning : Height UP/DOWN, Trendlenburg/Reverse Trendlenburg , Lateral tilt sides, Flex/Reflex, Chair, Kidney elevation through drive handle. , Base and column covered with S.S. Top to ensure high durability and hygiene for the operation room. , Mobile on Smooth running castors. , Floor lock facility.</p>

Note:-

1. Rate of each item must be quoted, otherwise the bid may be liable to be rejected.
2. Amount must be mentioned in, up to 2 decimal points (rounded)
3. Rate quoted for items must be inclusive of all taxes, no any Levy or will be paid separately after the quoted rate.
4. The committee through consensus can accept modification in the specification of above items, if any.

➤ अन्य शर्तें:-

1. कार्यादेश में वर्णित, निर्धारित समय-सीमा के अन्दर, सामग्री/उपकरणों की आपूर्ति करनी होगी, अन्यथा अग्रधन की राशि जब्त करते हुए नियमानुसार कानूनी कार्रवाई की जाएगी।
2. बिना किसी कारण बताये निविदा को रद्द करने का अधिकार अधोहस्ताक्षरी को सुरक्षित होगा।
3. किसी भी तरह के विवाद की स्थिति में न्यायिक क्षेत्र कैमूर (भभुआ) होगा।

निविदा से संबंधित विशेष जानकारी कैमूर जिले के वेबसाईट www.kaimur.bih.nic.in पर भी देखा जा सकता है। अन्य जानकारी हेतु जिला योजना कार्यालय, कैमूर (भभुआ) में कार्य दिवस/अवधि में भी सम्पर्क किया जा सकता है।

नोट :- निविदा प्रकाशन हेतु जिला पदाधिकारी, कैमूर (भभुआ) का अनुमोदन प्राप्त है।

4/08/06.21
उप विकास आयुक्त
कैमूर (भभुआ)